



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

On

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/206,329	12/08/1998	GENG ZHANG	970663.ORI	5359
7590 09/14/2004 NIKOLAI MERSEREAU AND DIETZ 820 INTERNATIONAL CENTRE 900 SECOND AVENUE SOUTH MINNEAPOLIS, MN 554023325			EXAMINER EVANISKO, GEORGE ROBERT	
			ART UNIT 3762	PAPER NUMBER

DATE MAILED: 09/14/2004

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

09/206,329

Applicant(s)

ZHANG ET AL.

Examiner

George R Evanisko

Art Unit

3762

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED. (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 6/22/04.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1,3-19 and 21-36 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1, 3-19, 21-36 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- ☐ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____
- ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____
- ☐ Notice of Informal Patent Application (PTO-152)
- ☐ Other: _____

DETAILED ACTION

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

Claims 1, 3-19, and 21-36 are rejected under 35 U.S.C. 103(a) as being unpatentable over Haefner et al.

Haefner meets the limitations of the term "reduced capacitance". In addition, Haefner describes figure 1 as a dual chamber cardioverter/defibrillator with pacing capability (col 3, ln 45) being able to operate in the unipolar or bipolar sensing modes having atrial and ventricular tip and ring electrodes, a can indifferent electrode, and a metallic housing. Haefner further describes in figure 2 the use of a system in the bipolar mode using the tip and ring electrodes and indifferent can electrode and meets the electrode combination limitations of claims 3, 4, 6, 11, 15, 21, 22, 24, 29, and 33.

Art Unit: 3762

Haefner et al disclose the claimed invention except for the other different combinations of unipolar or bipolar sensing between atrial electrodes (tip and/or ring), ventricular electrodes (tip and/or ring), and case/can electrodes, for a system to sense between all combinations of any two electrodes. It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the implantable electrical therapy device using unipolar and bipolar sensing as taught by Haefner with different combinations of unipolar or bipolar sensing between atrial electrodes (tip and/or ring), ventricular electrodes (tip and/or ring), and case/can electrodes for a system to sense between all combinations of any two electrodes since it was known in the art that implantable heart sensing systems (pacemakers, defibrillators, etc) use unipolar or bipolar sensing between atrial electrodes, ventricular electrodes, and case/can electrodes for a system to sense between all combinations of any two electrodes to provide a pacemaker/defibrillator/ICD system that can sense heart activity using any of the electrodes so that particular configurations can be chosen depending on the implantation of the leads, the desired sensing of a particular heart condition(s) or area, and/or to allow for the selection of an optimum sensing configuration.

Response to Arguments

Applicant's arguments filed 6/22/04 have been fully considered but they are not persuasive. The argument that the examiner's conclusion of obviousness is based upon improper hindsight reasoning is not persuasive since it must be recognized that any judgment on obviousness is in a sense necessarily a reconstruction based upon hindsight reasoning. But so long as it takes into account only knowledge which was within the level of ordinary skill at the time the claimed invention was made, and does not include knowledge gleaned only from the

Art Unit: 3762

applicant's disclosure, such a reconstruction is proper. See *In re McLaughlin*, 443 F.2d 1392, 170 USPQ 209 (CCPA 1971). Four prior art references (Sholder, Thompson, Wittkamp, or Greeninger) have been previously cited showing that it is obvious to one having ordinary skill in the art to have an implantable system use any electrode combination and therefore the rejection takes into account only knowledge which was within the level of ordinary skill at the time the claimed invention was made, and does not include knowledge gleaned only from the applicant's disclosure. The references provide motivation/incentive for the combination for the 103 rejection to provide additional sensing between additional electrodes used in pacemakers, defibrillators, and/or ICD's to sense heart activity using any combination of the electrodes so that particular configurations can be chosen depending on the implantation of the leads, the sensing of a particular heart condition(s) or area, and/or to allow for the selection of an optimum sensing configuration.

Conclusion

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event,

Art Unit: 3762

however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to George R Evanisko whose telephone number is 703 308-2612. The examiner can normally be reached on M-F 6:30-5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Angela Sykes can be reached on 703 308-5181. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

George R Evanisko
Primary Examiner
Art Unit 3762

9/10/4

GRE
September 10, 2004